

AESTHETIC ORTHODONTIC ARCHWIRES: THE STATE OF ART

M. TOTINO¹, A. RICCIO¹, M. LEO¹, L. ARCURI¹, L. CERRONI², G. PASQUANTONIO²,
R. CONDO²

¹ PhD in Materials for Health, Environment and Energy, University of Rome "Tor Vergata", Rome, Italy

² Department of Clinical Sciences and Translational Medicine, University of Rome "Tor Vergata", Rome, Italy

SUMMARY

Despite the improved aesthetics of this type of archwires, there are differing opinions in the literature about their actual clinical validity because of the problems that may arise during their use in the oral cavity. A systematic review was performed to analyze studies published over the last 25 years about aesthetic orthodontic archwires, in order to: analyze the technical characteristics and biotechnological properties, highlight the reliability and long-term effectiveness, and establish which are the most significant physical variables in determining the clinical efficacy during the orthodontic treatment and hypothesize future developments. A total sample of 24 articles was selected. Although initially the outer coating improves the aesthetic value of the orthodontic archwire, the wear that characterizes clinical use creates significant changes in surface, mechanical properties, such as sliding and friction, impacting negatively on the biocompatibility and aesthetics.

Thermoplastic polymers could represent an excellent alternative and a good future prospect for the production of orthodontic wires characterized by an high aesthetic value and optimum mechanical properties.

Key words: coated archwires, aesthetic archwires.

Introduction

In the last decade, the number of adult patients willing to face orthodontic treatment is increasing rapidly and, consequently, also the demand for orthodontic devices with high aesthetic value. It was therefore necessary to develop innovative materials that meet the expectation of adequate clinical performance, taking into account the aesthetic parameters of the patient (1).

To meet these demands, the industries have produced materials with low visual impact, such as aesthetic orthodontic brackets and aesthetic conventional ligatures (ring shape) in latex or polyurethane and unconventional aesthetic ligatures (Slide low Friction - Leone® Spa) in polyurethane (2-4).

In particular, as regards the construction of aes-

thetic orthodontic brackets, it has involved a wide range of materials, such as polycarbonates, ceramics (polycrystalline and single crystal aluminum oxide) and bioceramic (zirconium dioxide) (5).

The first aesthetic orthodontic archwire produced without metallic material, contained a core of silica, an intermediate layer of silicone resin and an outer layer of nylon (Optiflex Ormco) (6, 7).

Subsequently Fallis and Kusy devised a new aesthetic orthodontic archwire containing glass fibers (Owens Corning, Toledo, Ohio) embedded in a polymeric matrix composed of bisfenoloA-diglycidylether methacrylate and triethylene glycidimetacrilato (8).

Imai et al. devised an arch reinforced polymer fibers consisting of polymethylmethacrylate (PMMA) and $\text{CaO-P}_{(2)}\text{O}_{(5)}\text{-SiO}_{(2)}\text{-Al}_{(2)}\text{O}_{(3)}$ (CPSA) with the aim of further improving the aesthetic appearance, but also to make the mechanical properties

more similar to those of metal archwires (9).

Despite the excellent aesthetic characteristics of the latter, to date the coated metallic archwires are the only aesthetic orthodontic wires available for clinical use due to their own physical abilities of wear resistance and biocompatibility (10, 11). Of these, the most used in orthodontics are constituted by a core of NiTi alloy, coated with plastic resin materials (such as synthetic resin or fluorine-containing epoxy resin mainly composed of polytetrafluoroethylene or Teflon®) and hydrogenated carbon or zirconium dioxide resins (12, 13).

However, the best results are obtained with Teflon® coating NiTi archwires, which produces less friction than the corresponding uncoated (14).

Subsequently, were introduced other types of coating such as aluminum nitride and titanium (TiAlN), tungsten carbide/carbon (WC/C) and diamond-like carbon in order to protect the alloy from the corrosive effects of mouthwashes and toothpastes fluoride gel-based commonly recommended to orthodontic patient in the practice of oral care at home (15-17). Such of these coatings have limited corrosion phenomena and an average reduction of nickel ions released in the oral cavity of about 85% (11, 16, 17).

Have recently been introduced two types of NiTi aesthetic archwires, the Woowa (Dany Harvest, Seoul, South Korea) and the Bioforce High aesthetic (Dentsply GAC, Islandia, NY), whose surface coatings show enhanced properties for both aesthetic impact, which for the physical-mechanical characteristics, with a significant reduction of surface roughness compared to the untreated archwires, and a substantial increase in smoothness during the orthodontic movement, thereby ensuring a better biocompatibility of the final product (18).

The aim of this study was to perform a systematic review of the literature analyzing all the scientific articles published in recent years about aesthetic orthodontic archwires, in order to: analyze the technical characteristics and biotechnological properties, highlight the reliability and long-term effectiveness, and establish which are the most significant physical variables in determining the clinical efficacy during the orthodontic treatment and hypothesize future developments.

Materials and methods

To perform a systematic review of the literature were chosen initially some basic parameters, inclusion and exclusion criteria, which each article would have to meet in order to be considered appropriate to the purposes of this research.

It was conducted in MEDLINE and PUBMED database by choosing the following keywords: “coated archwires” and “aesthetic archwires” within the limits of research all inclusion and exclusion criteria described below.

Were analyzed studies published over the last 25 years, assuming a total sample of 24 articles.

The selected languages are English and Italian in order to ensure a clear understanding of the text.

Were evaluated scientific articles and reviews on experimental work carried out *in vivo* and *in vitro*.

This decision was taken in line with the objectives of this research, aimed to the study and understanding of the clinical efficacy of long-term orthodontic arches.

According to further inclusion and exclusion criteria chosen, were eliminated all the articles written in languages other than Italian and English, published before 1990, making use of experimental tests on animals.

According to these criteria were available for an initial analysis of 44 articles.

Of these, 11 were discarded because not congruent to title to the objectives and parameters established in the present research.

Subsequently other 9 of them were discarded by abstract because no respectful inclusion criteria chosen.

The remaining 24 articles were analyzed and divided into 3 groups.

The data collected were used to extrapolate general information on the use and effectiveness of aesthetic orthodontic archwires during orthodontic treatment. The purpose of this final analysis was to examine specific clinical parameters of aesthetic orthodontic archwires: the aesthetic factor and the clinical performance of the coating materials, the effectiveness of these materials on the basis of parameters such as mechanical properties of the load and deflection, corrosion resistance, and sliding

properties, and the reliability of the techniques of application of the materials themselves.

Results

The sample of articles has been divided into 3 groups according to the characteristics and the properties of each type of orthodontic archwires: in Group 1 (G_1) are included articles that examine the mechanical properties of load and deflection by means of certain tools such as the three-point bending test (6, 18, 19, 26-29);

in Group 2 (G_2) articles that evaluate the corrosion resistance (17, 30-32, 34-36);

while in Group 3 (G_3) than those who analyze the properties of the mechanical sliding (10, 13, 14, 40-42, 45), surface roughness and adhesion of the coating (18, 31, 32, 46-48).

Every aspect was analyzed individually and were assessed both changes that the veneering brings to the properties of conventional orthodontic archwires, both the modifications that the coatings themselves undergo as a result of their stay in the oral cavity by physical-chemical stress.

This is to assess the actual clinical efficacy and long-term therapeutic of such archwires.

Discussion

G1: Mechanical properties of load and deflection (3-point bending test)

The performance of an orthodontic arch depends on the material of which it is composed and the geometry of its cross-section.

The application process of coating includes a surface treatment of the wire and the use of Cathodic arc physical vapor deposition (PVD) or compressed air as a means of transport for the resin particles (polytetrafluoroethylene) sprayed to cover the archwire. The set wire-coating is then further heat-treated in a microwave cham-

ber (12, 13, 18).

The mechanical properties of the metal arches may be affected during this process and may be necessary any changes about the size of the inner core of the alloy to compensate for the thickness of the coating layer (19).

It's important to consider that, aesthetic archwires may have a metal core inside smaller sizes to compensate for the thickness of the coating layer. To achieve the same nominal size, the coated archwire is constructed from a smaller archwire that, when coated, reaches the nominal dimensions stated for the archwire (6).

Industries actually indicate as usual cross-sectional dimension of the aesthetic archwire, not the real size of the wire, but a value including the thickness of the set archwire-coating. Therefore, considering that the size of the metal wire is a substantial component of the resultant of the forces, certain variations of its section could affect the mechanical properties of the archwire in terms of torsion and rigidity and about its final clinical performance (20, 21).

These mechanical properties can be examined using a 3-point bending test. The advantage of a 3-point bending test is to be able to get a mechanical simulation very similar to clinical application comparing different orthodontic wires with super-elastic properties (22, 23). Through this test it's possible to analyze several parameters: the load curve, is the force required to engage the wire in the brackets; the discharge curve, which represents the force transferred to the teeth during the orthodontic treatment (24); the modulus of elasticity, which corresponds to the archwire stiffness or rigidity of the material (higher values indicate more rigid wires); the form of resiliency, or the ability of a material to store energy when it is deformed (25).

Starting from these considerations and using the 3-point bending test several study evaluates the properties of load-deflection of the aesthetic archwires, providing information on the behavior of the wires when undergoing deformation and comparing them with conventional orthodontic archwires.

Da Silva et al. have analyzed 3 groups of coated archwires: coated orthodontic wires with a cross-

sectional dimension identical to that of uncoated wires control group, archwires with all the surfaces completely coated, wires of which had been coated only the vestibular surface. The values about the mechanical properties of the archwires after a period of clinical use of 21 days, was compared with those of the corresponding unused orthodontic wires of the control group. These wires, both new and used, thus producing lower forces of loading and unloading, possess a greater modulus of elasticity, a lower modulus of resilience and they also demonstrate a minimum values of bending compared to the corresponding archwires of the control group (19).

A similar investigation into the mechanical and aesthetic properties of new generation coated nickel-titanium wires in the as-received state and after clinical use was conducted by Bradley et al. Sixty one subjects were randomly assigned to four groups, two groups of coated wires and two groups of comparable, non-coated controls. The period in the mouth ranged from 4 to 12 weeks after insertion. Three-point bending tests were done on as-received and used wires and its results indicate a wide variation in test results with large standard deviations among all the groups (26).

Elayyan et al. performed a randomized clinical trial about NiTi wires coated with epoxy resin. Run the 3-points bending test reported that the coated wire used has values of force loading and unloading lower than those typical of a new wire (27).

In a second study conducted in 2010 by Elayyann et al., four types of orthodontic archwires were investigated, 2 superelastic nickel-titanium and 2 coated Ultra-aesthetic archwires in 0.016-in and 0.018 x 0.025-in dimensions. All specimens were tested in a universal testing machine in a 3-point bending test. Coated superelastic wires produced statistically significantly lower forces in loading and unloading when compared with the superelastic nickel-titanium wires at most archwire deflections ($P < 0.01$). For all wires, an increase in size resulted in an increase in force (6).

Alavi in 2012 compares the load-deflection and surface properties of 3 types of archwires including ultra-aesthetic polycoated, ultra-aesthetic epoxyresin coated and conventional (uncoated) superelastic nickel-titanium (NiTi) archwires in con-

ventional and metal-insert ceramic brackets. They were tested in three-bracket bending test machine. It results epoxyresin-coated archwire had the lowest force and highest end load deflection point (ELDP) (28).

In orthodontics, have been recently introduced 2 new aesthetic NiTi wires, their trade names are: Woowa (Dany Harvest, Seoul, South Korea) and BioForce High Aesthetic (Dentsply GAC, Islandia, NY). Woowa has a structure of double-layer coating: the inner one is made of silver and platinum, instead the outer layer consist of highly aesthetic polymeric coating. Bioforce High Aesthetic is made of a coating in rhodium with a low reflectivity for the white color. About them, 3-points bending tests were carried out in order to: study the mechanical properties of these archwires, which were then compared with those of the not covered wires. The results showed that, in both cases, the coating adversely affects the mechanical behavior of the archwires (18, 19).

Therefore the reduction of the internal dimensions of the metallic core seems to be actually the variable most responsible for changes in the mechanical properties of the coated orthodontic wires (19). This is the reason why the latest development in aesthetic orthodontic appliances is a polymeric orthodontic archwire without metal inner core, with high springback and hight ductility. In a study performed in 2014 Varela et al. showed that mechanical strength of a new thermoplastic polymer archwire was lower than metallic alloys but is adequate for the first stage in orthodontic therapy (29).

G2: Corrosion resistance

In 1999 Kim analyzed significant differences in corrosion phenomena of stainless steel, nickel titanium, nitride and epoxy-coated nickel titanium and titanium orthodontic wires with the result that epoxy-coated nickel titanium wires exhibited the least corrosive potential (30).

Da Silva et al. have examined 4 types of aesthetic rectangular wires to evaluate the thickness, surface characteristics and stability of the outer coat-

ing before and after 21 days of exposure in the oral cavity, compared them with those belonged to uncoated stainless steel and NiTi archwires. The results of this search demonstrated that the aesthetic wires does not show a uniform thickness of the coating. These wires therefore have a lower aesthetic value in that they have a non-durable coating after exposure in the oral cavity. The remaining coating has also produced a greater surface roughness of the archwires examined than those in the control group (31, 32).

Cai et al. led a study on the corrosion resistance and fracture toughness of different types of archwires coated in comparison to traditional wires and scored three categories of arch: wire without modification, wire with crystallographic modifications and wire with Teflon coating out-distanced (33).

It's clear therefore that, the aesthetic coating preserves the orthodontic wire metal core by the corrosion processes, or by fluoride-induced corrosion and improve orthodontic friction (17, 34, 35) but the type and nature of coating material can effectively influence the anticorrosive features of NiTi wires, compared with its surface roughness values (36) and after a prolonged use in the oral cavity anyway, are themselves subject to corrosion processes. It depends on the action of the oral fluid, the nature of the material used for the cladding and the presence of imperfections due to an incorrect procedure of affixing the material itself (12). The corrosion of the archwire coating after clinical use in the oral cavity can also lead to undesirable results such as tartar buildup, microstructural and chemicals changes that cause a mechanical deformation of the arch and the release of chemical elements within the oral fluids (14).

G3: Properties of the mechanical sliding, surface roughness and adhesion of the coating

Therefore, the corrosion phenomena of the coating have negative impacts on biocompatibility, aes-

thetics and also adversely affecting the physical and mechanical properties such as the sliding ability of the wire (34).

The increase in friction due to the surface roughness is widely discussed in the literature (37-39). Recent studies have revealed a lower resistance to sliding of the orthodontic wires coated than uncoated (10, 13, 40, 41).

There are many different methods and materials used to improve the surface of the wire (16, 42-44). Farronato et al. in 2012 have made a study *in vitro* comparing some of those, and found that all coatings improve the sliding characteristics compared to those non-coated, but that the best results are obtained with the Teflon coating (14).

However, studies performed *in vivo* have shown that the clinical use of orthodontic wires coated increases their surface roughness and the level of friction (18, 28, 45-47).

Beside intra-oral exposure and archwire-bracket friction of coated wire altered the morphology and changed the elemental composition of the surface due to the process of corrosion, adhesion of organic matters and ionic exchange with oral fluids (48).

With the intention to overcome all these problems, the latest technological development in orthodontic research are projected into a solution regarding high aesthetic value that maintained contextually a satisfactory clinical efficacy. It has led to the production of a translucent polymer archwire with a considerable elastic return and a high ductility (49).

Marroco and Goldberg have developed a wire made of a new polymer based on polyphenylene, whose extremely rigid molecular structure bears a high load force in addition to possessing a high modulus of elasticity (50, 51).

The clinical trial has demonstrated the effectiveness of tooth movement with such wire during the first stage of orthodontic treatment (leveling and alignment) (52). Recently, Hadjichristidis et al. examined some copolymers: multigraft and polystyrene-polyisoprene-g, thermoplastic elastomers with high tensile strength and rupture (percentage increase of about 1550%) (53, 54).

The mechanical properties of these copolymers are guaranteed by the architecture and morpholo-

gy of both molecules in fact, these polymers have a very similar behavior to superelastic materials. From this it derives an aesthetic archwire that has therefore excellent mechanical properties and good plasticity, parameters which allow to perform a similar curvature to that of stainless steel orthodontic wires. Furthermore, the coefficient of friction is lower than that of uncoated metal ones (55, 56).

Finally, an interesting study in 2014 conducted by Valera et al. introduces a new thermoplastic polymer for orthodontic applications, obtained and extruded into wires with round and rectangular cross sections. Tensile, 3- point bending test, friction and stress relaxation behaviour, and formability characteristics were assessed. The results have demonstrated that stresses delivered were generally slightly lower than typical beta-titanium and nickel-titanium archwires. The polymer wire has good instantaneous mechanical properties; tensile stress decayed about 2% over 2 hours depending on the initial stress relaxation for up to 120 hours. High formability allowed shape bending similar to that associated with stainless steel wires. The friction coefficients were lower than the metallic conventional archwires improving the slipping with the brackets (57). In conclusion, according the Authors this new polymer could be a good candidate for aesthetic orthodontic archwires (29, 58).

Conclusions

In conclusion, despite the improved aesthetics of this type of archwires, there are differing opinions in the literature about their actual clinical validity because of the problems that may arise during their use in the oral cavity.

Although initially the outer coating improves the aesthetic value of the orthodontic archwire, the wear that characterizes clinical use creates significant changes in surface, mechanical properties, such as sliding and friction, impacting negatively on the biocompatibility and aesthetics (47, 58).

Thermoplastic polymers could represent an excellent alternative and a good future prospect for the production of orthodontic wires characterize by an

high aesthetic value and optimum mechanical properties (29, 49-55).

References

1. Russell JS. Aesthetic orthodontic brackets. *J Orthod.* 2005;32:146-63.
2. Baccetti T, Franchi L, Camporesi M. Forces in the presence of ceramic versus stainless steel brackets with unconventional vs conventional ligatures. *Angle Orthod.* 2008;78:120.
3. Condò R, Casaglia A, Armellini E, Condò SG, Cerroni L. Traditional elastic ligatures versus slide ligation system. A morphological evaluation. *Oral Implantol (Rome)*. 2013 Jul 15;6(1):15-24. eCollection 2013.
4. Condò R, Casaglia A, Condò SG, Cerroni L. Plaque retention on elastomeric ligatures. An in vivo study. *Oral Implantol (Rome)*. 2013 Mar 19;5(4):92-9.
5. Condò R, Casaglia A, Cozza P. SEM analysis of zirconium brackets using MIM technology. *Minerva Stomatol.* 2005 Apr;54(4):207-17.
6. Elayyan F, Silikas N, Bearne D. Mechanical properties of coated superelastic archwires in conventional and self-ligating orthodontic brackets. *Am J Orthod Dentofacial Orthop.* 2010;137:213-7.
7. Talass MF. Optiflex archwire treatment of a skeletal class III open bite. *J Clin Orthod.* 1992;26:245-252.
8. Fallis DW, Kusy RP. Variation in flexural properties of photopultruded composite archwires: analyses of round and rectangular profiles. *J Mater Sci Mater Med.* 2000;11:683-693.
9. Imai T, Watari F, Yamagata S, Kobayashi M, Nagayama K, Nakamura S. Effects of water immersion on mechanical properties of new esthetic orthodontic wire. *Am J Orthod Dentofacial Orthop.* 1999;116:533-8.
10. Zufall SW, Kusy RP. Sliding mechanics of coated composite wires and the development of an engineering model for binding. *Angle Orthod.* 2000;70:34-47.
11. Kusy RP. Orthodontic biomaterials: From the past to the present. *Angle Orthod.* 2002;72:501-2.
12. Krishnan V, Krishnan A, Remya R, Ravikumar KK, Asha Nair S, Shibli SMA, Varma HK, Sukumaran K, Jyothindra Kumar K. Development and evaluation of two PVD-coated b-titanium orthodontic archwires for fluoride-induced corrosion protection. *Acta Biomaterialia.* 2011;7:1913-1927.
13. Husmann P, Bourauel C, Wessinger M, Jäger A. The frictional behavior of coated guiding archwires. *J Orofac Orthop.* 2002;63(3):199-211.
14. Farronato G, Maijer R, Caria MP, Esposito L, Alberzoni D, Cacciatore G. The effect of Teflon coating on the resistance to sliding of orthodontic archwires. *European Journal of Orthodontics.* 2012;34:410-417.
15. Walker MP, White RJ, Kula KS. Effect of fluoride pro-

- phylactic agents on the mechanical properties of nickel-titanium-based orthodontic wires. *Am J Orthod Dentofacial Orthop.* 2005;127:662-9.
16. Ohgoe Y, Hirakuri KK, Ozeki K, Fukui Y. Investigation of diamond-like carbon coating for orthodontic archwire, New Diamond and Frontier Carbon Technology. 2007;17(6):281-288.
17. Huang SY1, Huang JJ, Kang T, Diao DF, Duan YZ. Coating NiTi archwires with diamond-like carbon films: reducing fluoride-induced corrosion and improving frictional properties. *J Mater Sci Mater Med.* 2013 Oct;24(10):2287-92.
18. Iijimaa M, Muguruma T, Brantley C, Choed H, Nakagaki S, Alapatif S, Mizoguchi I. Effect of coating on properties of esthetic orthodontic nickel-titanium wires. *Angle Orthodontist.* 2012;82(2):319-325.
19. da Silva DL, Mattos CT, Franzotti Sant'Anna E, de Oliveira Ruellas AC, Elias CN. Cross-section dimensions and mechanical properties of esthetic orthodontic coated archwires. *American Journal of Orthodontics and Dentofacial Orthopedics.* 2013 April;143(4)-Supplement 1:85-91.
20. Meling T, Ødegaard J, Meling E. On mechanical properties of square and rectangular stainless steel wires tested in torsion. *Am J Orthod Dentofacial Orthop.* 1997;111:310-20.
21. Meling T, Ødegaard J. On the variability of cross-sectional dimensions and torsional properties of rectangular nickel-titanium arch wires. *Am J Orthod Dentofacial Orthop.* 1998;113:546-57.
22. Krishnan V, Kumar KJ. Mechanical properties and surface characteristics of three archwire alloys. *Angle Orthod.* 2004 Aug;74(4):533-8.
23. Walker M, White R, Kula K. Effect of fluoride prophylactic agents on the mechanical properties of nickel-titanium-based orthodontic wires. *Am J Orthod Dentofacial Orthop.* 2005;127:662-9.
24. Segner D, Ibe D. Properties of superelastic wires and their relevance to orthodontic treatment. *Eur J Orthod.* 1995;17:395-402.
25. Brantley WA. Orthodontic wires. In: Brantley WA, Eliades T, editors. *Orthodontic materials: scientific and clinical aspects.* Stuttgart, Germany, and New York: Thieme; 2001. p. 77-103.
26. Bradley TG, Berzins DW, Valeri N, Pruszyński J, Eliades T, Katsaros C. An investigation into the mechanical and esthetic properties of new generation coated nickel-titanium wires in the as-received state and after clinical use. *Eur J Orthod.* 2014 Jun;36(3):290-6.
27. Elayyan F, Silikas N, Bearn D. Ex vivo surface and mechanical properties of coated orthodontic archwires. *Eur J of Orthod.* 2008;30:661-667.
28. Alavi S, Hosseini N. Load-deflection and surface properties of coated and conventional superelastic orthodontic archwires in conventional and metal-insert ceramic brackets. *Dent Res J (Isfahan).* 2012 Mar; 9(2): 133-8.
29. Varela JC, Velo M, Espinar E, Llamas JM, Rùperez E, Manero JM, Javier Gil F. Mechanical properties of a new thermoplastic polymer orthodontic archwire. *Materials Science and Engineering.* 2014 May;C 42:1-6.
30. Kim H, Johnson JW. Corrosion of stainless steel, nickel-titanium, coated nickel-titanium, and titanium orthodontic wires. *Angle Orthod.* 1999 Feb; 69(1):39-44.
31. da Silva DL, Mattos CT, de Araújo MV, de Oliveira Ruellas AC. Color stability and fluorescence of different orthodontic esthetic archwires. *Angle Orthod.* 2013;83:127-132.
32. da Silva DL, Mattos CT, Simão RA, de Oliveira Ruellas AC. Coating stability and surface characteristics of esthetic orthodontic coated archwires. *Angle Orthod.* 2013 Nov;83(6):994-1001.
33. Cai F, Yang Q, Huang X, Wei R. Microstructure and corrosion behavior of CrN and CrSiCN coatings. *J Mat EngPerform.* 2010;19(5):721-727.
34. New Neumann P, Bourauel C, Jäger A. Corrosion and permanent fracture resistance of coated and conventional orthodontic wires. *J Mater Sci Mater Med.* 2002;13(2):141-147.
35. Kobayashi S, Ohgoe Y, Ozeki K, Hirakuri K, Aoki H. Dissolution effect and cytotoxicity of diamond-like carbon coatings on orthodontic archwire. *J Mater Sci Mater Med.* 2007 Dec;18(12):2263-8.
36. Krishnan M, Seema S, Kumar AV, Varthini NP, Sukumaran K, Pawar VR, Arora V. Corrosion resistance of surface modified nickel titanium arch wires. *Angle Orthod.* 2014 Mar; 84(2):358-67.
37. Kusy RP, Whitley JQ, Mayhew MJ, Buckthal JE. Surface roughness of orthodontic archwires via laser spectroscopy. *Angle Orthod.* 1988;58:33-45.
38. Prosser RR, Bagby MD, Erickson LC. Friction and roughness of nickel-titanium arch wires. *Am J Orthod Dentofacial Orthop.* 1991;100:341-38.
39. Tselepis M, Brockhurst P, West VC. Frictional between brackets and archwires. *Am J Orthod Dentofacial Orthop.* 1994;106:131-13.
40. Muguruma T, Iijima I, Brantley WA, Mizoguchi I. Effects of a diamond-like carbon coating on the frictional properties of orthodontic wires. *Angle Orthod.* 2011;81(1):141-148.
41. Doshi UH, Bhad-Patil WA. Static frictional force and surface roughness of various bracket and wire combinations. *Am J Orthod Dentofacial Orthop.* 2011;139(1):74-79.
42. Redlich M, Katz A, Rapoport L, Wagner HD, Feldman Y, Tenne R. Improved orthodontic stainless steel wires coated with inorganic fullerene-like nanoparticles of WS(2) impregnated in electroless nickelphosphorous Im. *Dental Materials.* 2008;24:1640-1646.
43. Kusy RP, Saunders CR, Whitley JQ. Improving arch mechanics through surface chemistry. In: Nanda R (ed). *Biomechanics in clinical orthodontics.* WB Saunders Company, Philadelphia 1997; pp. 50-6.

44. Kusy RP, Keith O, Whitley JQ, Saunders CR. Coefficient of friction characterization of surface-modified polycrystalline alumina. *Journal of the American Ceramic Society*. 1993;76:336-342.
45. Wichelhaus A, Geserick A, Hibst R, Sander FG. The effect of surface treatment and clinical use on friction in NiTi orthodontic wires. *Dent Mater*. 2005;21:938-945.
46. D'Antò V, Rongo R, Ametrano G, et al. Evaluation of surface roughness of orthodontic wires by means of atomic force microscopy. *Angle Orthod*. 2012;82:922-928.
47. Rongo R, Ametrano G, Gloria A, Spagnuolo G, Galeotti A, Paduano S, Valletta R, D'Antò V. Effects of intraoral aging on surface properties of coated nickel-titanium archwires. *Angle Orthod*. 2014 Jul;84(4):665-72.
48. Zegan G, Sodor A, Munteanu C. Surface characteristics of retrieved coated and nickel-titanium orthodontic archwires. *Rom J Morphol Embryol*. 2012;53(4):935-9.
49. Goldberg AJ, Liebler AH, Burstone CJ. Viscoelastic properties of an aesthetic translucent orthodontic wires. *Eur J Orthod*. 2011 Dec;33(6):673-678.
50. Marrocco ML, Gagne RR, Trimmer MS. Rigid rod polymers. 2000. US patent 6, 087, 467.
51. Goldberg AJ, Burstone CJ. Advanced thermoplastics for orthodontics 2007, US patent 7, 186, 115.
52. Kuhlberg AJ. Esthetic orthodontic wire clinical trial. *New Ortho Polymers*. Farmington, 2009.
53. Arciniegas M, Manero JM, Peña J, Gil FJ, Planel JA. Study of new multifunctional shape memory and low elastic modulus Ni-free Ti alloys. *Mater Sci Eng*. 2008 Dec; A 39:742-751.
54. Duan Y, Rettler E, Schneider K, Schlegel R, Thunga M, Weidisch R, Siesler H.W, Stam M, Mays J.N, Hadjristidis N. Deformation behaviour of sphere forming trifunctional multigraft copolymer. *Macromolecules*. 2008;42:4565-4568.
55. Schlegel R, Wilkin D, Duan R, Weidisch Y, Heinrich G, Uhrig D, Mays JN, Jatrou H, Hadjristidis N. Stress softening of multigraft copolymer. *Polymer*. 2009;50: 6297-6304.
56. Gargari M, Gloria F, Napoli E, Pujia AM. Zirconia: cementation of prosthetic restorations. Literature review. *Oral Implantol (Rome)*. 2011 Jan 23;3(4):25-9.
57. Gargari M, Gloria F, Cappello A, Ottria L. Strength of zirconia fixed partial dentures: review of the literature. *Oral Implantol (Rome)*. 2011 Jan 23;3(4):15-24.
58. Kaneko K, Yokoyama K, Moryama K, Asaoka K, Sakai J. Degradation in performances of orthodontic wires caused by hydrogen absorption during short time immersion in 2.0% acidulate phosphate fluoride solution. *Angle Orthod*. 2004;74:487-495.

Correspondence to:

Dr.ssa Adele Riccio
Dipartimento di Ortognatodonzia
Centro Odontostomatologico "G.B. Orsenigo"
"Ospedale San Pietro Fatebenefratelli", Roma, Italy
E-mail: adele.riccio@gmail.com